

The Lodge 37 Barnett Way Barnwood Gloucester GL4 3RT Tel: +44(0)1452 632712 sales@thesempregroup.com

INDUSTRIAL METROLOGY | MEASURING INSTRUMENTS | LEAN ENTERPRISE | ENGINEERING SERVICES

Form CRS172- Issue 1 - 06/12/2021

Uncontrolled When Printed

Page 1 of 1

## **Calibration Certificate**

For Opticline Systems

Certificate Number -

MDS-07334-B9L7

Stourbridge Turning & Grinding Rufford Road

Stourbridge DY9 7NE

This The Sempre Group Calibration Certificate documents the function of components in the Opticline system using the calibrated values of a reference standard. Any new or repaired components were set and checked prior to the issue of this certificate.

Measurement conditions under which the certificate was issued conform to the pertinent ISO regulations. The standards utilised are subject to gauge management criteria.

The user is responsible for keeping an appropriate recalibration schedule; we recommend twelve-monthly checks.

**EQUIPMENT** Measurment Uncertainty D=2+D/100 L=5+L/100 (Result in  $\mu$ )

Instrument Type	Hommel Opticline		Automatic Profile Measuring instrument		
Serial Number	10200013				
Software Version		3.71.2	22.0		

## **STANDARD**

<u>Standard Type</u>	<u>Calibration Number</u>	Calibration Date of Standard
MET132_L108(1)	Ø2-12 /L108-015028-312-51- 56655-D-K-15151-01-00-2023-03	17 March 2023
MET134_L165(1)	Ø8-140/L165 10021530/41191- D-K-15151-01-00-2021-04	16th April 2021
MET133_L211(1)	Ø8-140/L211 10021525/014 - 56656- D-K-15151-01-00-2023-	17 March 2023

Standard calibrated at DKD Accredited Calibration Laboratory. The uncertainty of the laboratory is (eumetron GmbH) - U = 0.3 μm + 0.6 • 10-6 • D; D is the diameter - U = 0,40 μm + 0,60 • 10-6 • L; L is the Length **MEASURING REPORTS** 

Max Diameter Error	0.000220	At	38MM	Pass/Fail	Pass
Max Length Error	0.000584	At	130MM	Pass/Fail	Pass

Inspected by	Lee Roberts	Approved by	Lee Roberts
Date	02/05/2024	Date	02/05/2024

